

Factsheet: Legionella Count Kits

Test Kit Specifications

Sample volume	100 mL	
Target cells	<i>Legionella spp.</i>	<i>L. pneumophila</i> SG 1-15 or SG 1
Time to results	First result in 1 hour 27 minutes. All 4 results in 2 hours.	First result in 1 hour 27 minutes. All 4 results in 2 hours.
Limit of detection (LOD)		
Drinking water	65 cells/100 mL (0-65 CFU/100 mL)	50 cells/100 mL (0-50 CFU/100 mL)
Process water	200 cells/100 mL (0-200 CFU/100 mL)	150 cells/100 mL (0-150 CFU/100 mL)
Sensitivity	100%	100%
Specificity	98%	98%
Quantitative working range		
Drinking water	200 - 5,000,000 cells/100 mL	150 - 5,000,000 cells/100 mL
Process water	600 - 5,000,000 cells/100 mL	500 - 5,000,000 cells/100 mL
Inclusivity list	<i>L. pneumophila</i> SG 1-15, <i>L. anisa</i> , <i>L. bozemanii</i> , <i>L. dumoffii</i> , <i>L. long-beachae</i> , <i>L. micdadei</i> , <i>L. parisiensis</i>	<i>L. pneumophila</i> SG 1-15 or <i>L. pneumophila</i> SG 1

Test Kit Facts

Unit of results	Number of viable cells per 100 mL
Main steps in the workflow	Filter 100 mL sample, resuspend bacteria in 1 mL buffer, add 200 µL to reaction tube, await 60 minutes, dilute, add sample to cartridge and start analysis
Daily throughput	Up to 28 samples per day per rqmicro.COUNT instrument
Main areas of application	Cooling tower monitoring, microbial control, wastewater reuse, applied research, water safety concept, outbreak management, Legionella risk assessment
Regulatory status of the test kit	The Association of German Engineers (VDI) included the method in their guideline for Legionella monitoring (42. BImSchV, VDI-Richtlinie 4250 Blatt 2).
Comparison rqmicro test results versus cultivation-based test results	Third-party lab validations confirmed high correlation of rqmicro results to cultivation-based tests according to ISO 11731:2017 in spiking experiments. In practice, a large and highly variable share of viable but non-culturable cells (VBNC) of Legionella remains undetected by cultivation-based methods but is included in the single-cell counting method used by rqmicro.
Comparison rqmicro test results versus PCR-based test results	Third-party lab validations confirmed high correlation of rqmicro results to PCR results in spiking experiments. In practice, the viability assessment of PCR-based tests is prone to failure and correlation to the rqmicro method is therefore lower.
Laboratory equipment required	Filtration unit, vacuum pump, vortex mixer, rotator/roller mixer, forceps, pipettes, sample bottle
Shelf life	1.5 years
Storage conditions	Cold storage at 4 - 8 °C for reagents (reaction tubes, buffers, positive control)

General Method Description

Technologies used in the rqmicro test method	Specific or unspecific labelling of bacterial cells with staining dyes and subsequent quantification of cells using cartridge-based flow cytometry. Depending on the test kit, the rqmicro method can include cell concentration (filtration), immunomagnetic separation of target cells and/or additional staining to detect membrane integrity.
Laboratory or on-site	Both is feasible using the rqmicro.COUNT instrument. Specific tests (Legionella, E. coli) require additional equipment for sample preparation.

Frequently Asked Questions

How do I check test results?	Test results are immediately available on the rqmicro.COUNT instrument and on rqmicro's Cloud Solution.
How do you assure the safety of data on the Cloud Solution?	Data produced by rqmicro.COUNT instruments is securely hosted in Europe using market-leading hosting services by Amazon Web Services (AWS). All software services provided by rqmicro are covered by the ISO 9001 quality management system.
How does the method differentiate between live and dead cells?	Flow cytometry is able to assess the viability of single cells at a very high throughput of up to 10,000 cells per second. Membrane-damaged, and therefore, dead cells produce an additional signal that enables a differentiation from viable cells with intact cell membranes.
Could rqmicro rapid test give a positive result whilst the culture test gives a negative result?	Yes, because the rqmicro test also quantifies viable but non-culturable cells (VBNC) which are not detected in cultivation-based methods. Results obtained with the single-cell counting technology deployed in rqmicro.COUNT instruments are therefore more precise and often result in higher cell numbers when compared to cultivation-based tests.
Is the process automated?	Sample preparation takes 1-5 minutes hands-on-time per sample, depending on the test (ICC, TCC, Legionella, E. coli). The analysis on rqmicro.COUNT is fully automated.
Does rqmicro provide training?	New rqmicro.COUNT customers typically receive a 1 day training on the instrument and benefit from continued support by rqmicro Application Specialists.
Does rqmicro provide service contracts?	The rqmicro.COUNT Service Contract offers instrument service, extended warranty, software packages and continued application support for recurring test kit customers.



rqmicro AG
Brandstrasse 24
8952 Schlieren
Switzerland

Visit our website:
www.rqmicro.com